

ADDENDUM NUMBER 2

DATED: APRIL 5, 2018

**TO SPECIFICATIONS FOR
REQUEST FOR PROPOSAL #31024
FOR
PATIENT MONITORING EQUIPMENT AND SUPPLIES
DATED: MARCH 16, 2018**

The above entitled specifications are hereby modified as follows and except as set forth herein otherwise remain unchanged and in full force and effect.

REVISED RESPONSES TO SUPPLIER QUESTIONS #35 & #36:

Cerner Interface / Integration

35. Please confirm MU Health Care's requirement that all vendor quotes include ALL costs associated with complete and functioning Outbound HL7 Interface for Vital Signs to Cerner EMR at the time of go-live for all (723) Devices

~~**RESPONSE:** Only Telemetry would require HL7 interface all hardwired monitors are integrated through exprezznet.~~

REVISED RESPONSE: Yes, full functional outbound integration to Cerner EMR is required for all devices that currently send data at time of go-live.

36. Please confirm MU Health Care's requirement for vendor quotes to include full functioning Inbound ADT interface for Cerner (patient demographic data) the time of go-live for all (723) Devices

~~**RESPONSE:** Only Telemetry would require ADT interface all hardwired monitors are integrated through exprezznet.~~

REVISED RESPONSE: Yes, quotes should include full functioning inbound ADT interface for Cerner, for all bedsides, at the time of go-live.

THE CURATORS OF THE
UNIVERSITY OF MISSOURI

By: Laura E. Roth, CPPB
Strategic Sourcing Specialist
University of Missouri System Supply Chain